Article

Engineered Embodiment: Comment on "The embodiment of assistive devices-from wheelchair to exoskeleton" by M. Pazzaglia & M. Molinari

Kannape, Oliver Alan and Lenggenhager, Bigna

Available at <http://clock.uclan.ac.uk/13562/>

Kannape, Oliver Alan ORCID: 0000-0002-2930-4118 and Lenggenhager, Bigna (2016) Engineered Embodiment: Comment on "The embodiment of assistive devices-from wheelchair to exoskeleton" by M. Pazzaglia & M. Molinari. Physics of Life Reviews, 16 . pp. 181-183. ISSN 1571-0645

It is advisable to refer to the publisher's version if you intend to cite from the work.

<http://dx.doi.org/10.1016/j.plrev.2016.01.011>

For more information about UCLan's research in this area go to <http://www.uclan.ac.uk/researchgroups/> and search for <name of research Group>.

For information about Research generally at UCLan please go to <http://www.uclan.ac.uk/research/>

All outputs in CLoK are protected by Intellectual Property Rights law, including Copyright law. Copyright, IPR and Moral Rights for the works on this site are retained by the individual authors and/or other copyright owners. Terms and conditions for use of this material are defined in the [policies](#) page.

Accepted Manuscript

Engineered embodiment

Oliver Alan Kannape, Bigna Lenggenhager

PII: S1571-0645(16)00012-9
DOI: <http://dx.doi.org/10.1016/j.plrev.2016.01.011>
Reference: PLREV 705

To appear in: *Physics of Life Reviews*

Received date: 12 January 2016

Accepted date: 21 January 2016

Please cite this article in press as: Kannape OA, Lenggenhager B. Engineered embodiment. *Phys Life Rev* (2016), <http://dx.doi.org/10.1016/j.plrev.2016.01.011>

This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

Engineered Embodiment:**Comment on “The embodiment of assistive devices-from wheelchair to exoskeleton” by M. Pazzaglia & M. Molinari**

Oliver Alan Kannape¹ & Bigna Lenggenhager²

¹ School of Psychology, University of Central Lancashire, Preston, UK

² Neuropsychology Unit, Department of Neurology, University Hospital Zurich, Zurich, Switzerland

Corresponding author:

Email address: okannape@uclan.ac.uk

Character count: 5125 / 5985 (without/with whitespaces)

From brain-computer interfaces to wearable robotics and bionic prostheses – intelligent assistive devices have already become indispensable in the therapy of people living with reduced sensorimotor functioning of their physical body, be it due to spinal cord injury, amputation or brain lesions [1]. Rapid technological advances will continue to fuel this field for years to come. As Pazzaglia and Molinari [2] rightly point out, progress in this domain should not solely be driven by engineering prowess, but utilize the increasing psychological and neuroscientific understanding of cortical body-representations and their plasticity [3]. We argue that a core concept for such an integrated embodiment framework was introduced with the formalization of the forward model for sensorimotor control [4]. The application of engineering concepts to human movement control paved the way for rigorous computational and neuroscientific analysis. The forward model has successfully been adapted to investigate principles underlying aspects of bodily awareness such as the sense of agency in the comparator framework [5]. At the example of recent advances in lower limb prostheses, we propose a cross-disciplinary, integrated embodiment framework to investigate the sense of agency and the related sense of body ownership for such devices. The main onus now is on the engineers and cognitive scientists to embed such an approach into the design of assistive technology and its evaluation battery.

The initial challenge for assistive devices aiding bipedal locomotion is to ensure balance and reliability; unlike for the majority of upper limb movements, a single misstep may have severe consequences. Locomotion is largely reflex-driven and reliant on spinal cord pattern generators; often, only aspects such as initiation and termination are considered truly volitional. In part for these reasons and in part due to technological limitations, commercially available prostheses are to this day largely restricted to passive devices that rely on spring or damping mechanisms offering little adaptability and increasing the amputees' energy expenditure [6]. Only recently have active, powered lower limb prostheses become commercially available [7,8]. While the most common control strategy for these prostheses are finite state machines based on abstracted gait cycles, the availability of microcontrollers has already brought forth biomimetic control strategies incorporating neuromuscular models [9]. These have been shown to normalize walking and improve speed adaptability [10] but more intriguingly lead to an important conceptual consideration – to what extent can we embody (cognition on) the prosthesis?

Besides improved control algorithms, active prostheses have opened up the possibility of hybrid and fully volitional controllers through neural interfaces. Amputees, at least in the laboratory, are now able to use their residual muscles to control ankle push-off [11,12], stiffness over the gait-cycle [13], and even the movement of a biarticulate prosthesis [14]. But, do such sophisticated interfaces increase the sense of agency over the prostheses? Is this reflected in an increased sense of ownership and a more complete feeling of embodiment? There is a clear need to systematically address these points and evaluate prosthesis performance over and above the calculation of transition work and joint-mechanics. The implications extend to general health themes as an increased embodiment of the prosthesis may lead to an improved own body image, a more active life style, and decreased social isolation [15].

Apart from the psychological effect on the amputee, an integrated embodiment framework may teach us about the consequences of an altered sensorimotor system and its effects on locomotor control, motor learning, and body-awareness. This in turn provides a unique possibility for a comparison with psychiatric or neurological conditions of altered sensorimotor processes and embodiment [16]. Evaluating prosthetic devices using this integrated framework may further bring to light compensatory strategies related to multisensory integration and a potential reweighting of efferent and re-afferent information – both of great importance to refine neuroscientific models and develop therapies for neurologically-caused disorders of embodiment. To address all these issues, it is our strong opinion that research is not limited to patient populations. Analogue to using virtual reality to introduce visuotactile and visuomotor conflicts in ownership [17,18] and agency research [19–21], assistive devices such as orthoses and exoskeletons should be employed to investigate the effect of altered sensorimotor feedback in healthy participants. The autonomous exoskeleton developed by Mooney et al. [22] would for example allow for systematically varying the external power input during ankle plantar-flexion and methodically documenting adaptive sensorimotor control and accompanying changes in body-awareness.

It is quite likely that the human brain is capable of developing distinct forward models [23] – adapted to the type of assistive device, the control strategy employed, and ultimately the reafferent sensory information available to the user (especially with the addition of bidirectional neural interfaces). Being able to systematically modulate the inputs and

reafferent signals to those models will not only allow us to further improve assistive technology and investigate principles of sensorimotor control but to potentially tease out a set of rules underlying embodiment, independent of the end-effector. Interestingly, current ethical debates largely focus on whether or not assistive devices should be used to not only replace lost functionality but augment it, transcending the limitations of the healthy human body [24,25]. An antecedent concern, at least from a neurobiological perspective, may be to determine *what*, not *who*, defines a healthy human body.

References

- [1] Craelius W. The Bionic Man: Restoring Mobility. *Science* 2002;295:1018–21. doi:10.1126/science.295.5557.1018.
- [2] Pazzaglia M, Molinari M. The embodiment of assistive devices—from wheelchair to exoskeleton. *Phys Life Rev* n.d. doi:10.1016/j.plrev.2015.11.006., this issue.
- [3] Blanke O. Multisensory brain mechanisms of bodily self-consciousness. *Nat Rev Neurosci* 2012;13:556–71. doi:10.1038/nrn3292.
- [4] Wolpert DM, Diedrichsen J, Flanagan JR. Principles of sensorimotor learning. *Nat Rev Neurosci* 2011;12:739–51. doi:10.1038/nrn3112.
- [5] Blakemore SJ, Wolpert DM, Frith CD. Abnormalities in the awareness of action. *Trends Cogn Sci* 2002;6:237–42.
- [6] Hsu M-J, Nielsen DH, Lin-Chan S-J, Shurr D. The effects of prosthetic foot design on physiologic measurements, self-selected walking velocity, and physical activity in people with transtibial amputation. *Arch Phys Med Rehabil* 2006;87:123–9. doi:10.1016/j.apmr.2005.07.310.
- [7] Au SK, Herr H, Weber J, Martinez-Villalpando EC. Powered ankle-foot prosthesis for the improvement of amputee ambulation. *Conf Proc Annu Int Conf IEEE Eng Med Biol Soc IEEE Eng Med Biol Soc Annu Conf* 2007;2007:3020–6. doi:10.1109/IEMBS.2007.4352965.
- [8] Bellmann M, Schmalz T, Blumentritt S. Comparative biomechanical analysis of current microprocessor-controlled prosthetic knee joints. *Arch Phys Med Rehabil* 2010;91:644–52. doi:10.1016/j.apmr.2009.12.014.
- [9] Markowitz J, Krishnaswamy P, Eilenberg MF, Endo K, Barnhart C, Herr H. Speed adaptation in a powered transtibial prosthesis controlled with a neuromuscular model. *Philos Trans R Soc Lond B Biol Sci* 2011;366:1621–31. doi:10.1098/rstb.2010.0347.
- [10] Herr HM, Grabowski AM. Bionic ankle-foot prosthesis normalizes walking gait for persons with leg amputation. *Proc Biol Sci* 2012;279:457–64. doi:10.1098/rspb.2011.1194.
- [11] Wang J, Kannape OA, Herr HM. Proportional EMG control of ankle plantar flexion in a powered transtibial prosthesis. *IEEE Int Conf Rehabil Robot Proc* 2013;2013:6650391. doi:10.1109/ICORR.2013.6650391.
- [12] Kannape OA, Herr HM. Volitional control of ankle plantar flexion in a powered transtibial prosthesis during stair-ambulation. *Conf Proc Annu Int Conf IEEE Eng Med Biol Soc IEEE Eng Med Biol Soc Annu Conf* 2014;2014:1662–5. doi:10.1109/EMBC.2014.6943925.
- [13] Huang S, Wensman J, Ferris D. Locomotor Adaptation by Transtibial Amputees Walking with an Experimental Powered Prosthesis Under Continuous Myoelectric Control. *IEEE Trans Neural Syst Rehabil Eng Publ IEEE Eng Med Biol Soc* 2015. doi:10.1109/TNSRE.2015.2441061.
- [14] Hargrove LJ, Simon AM, Young AJ, Lipschutz RD, Finucane SB, Smith DG, et al. Robotic leg control with EMG decoding in an amputee with nerve transfers. *N Engl J Med* 2013;369:1237–42. doi:10.1056/NEJMoa1300126.
- [15] Sansam K, Neumann V, O'Connor R, Bhakta B. Predicting walking ability following lower limb amputation: a systematic review of the literature. *J Rehabil Med* 2009;41:593–603. doi:10.2340/16501977-0393.
- [16] Brugger P, Lenggenhager B. The bodily self and its disorders: neurological, psychological and social aspects. *Curr Opin Neurol* 2014;1. doi:10.1097/WCO.0000000000000151.

- [17] Lenggenhager B, Tadi T, Metzinger T, Blanke O. Video ergo sum: manipulating bodily self-consciousness. *Science* 2007;317:1096–9.
- [18] Blanke O, Slater M, Serino A. Behavioral, Neural, and Computational Principles of Bodily Self-Consciousness. *Neuron* 2015;88:145–66. doi:10.1016/j.neuron.2015.09.029.
- [19] Kannape OA, Blanke O. Agency, gait and self-consciousness. *Int J Psychophysiol Off J Int Organ Psychophysiol* 2012;83:191–9. doi:10.1016/j.ijpsycho.2011.12.006.
- [20] Kannape OA, Blanke O. Self in motion: sensorimotor and cognitive mechanisms in gait agency. *J Neurophysiol* 2013;110:1837–47. doi:10.1152/jn.01042.2012.
- [21] Kannape OA, Barré A, Aminian K, Blanke O. Cognitive Loading Affects Motor Awareness and Movement Kinematics but Not Locomotor Trajectories during Goal-Directed Walking in a Virtual Reality Environment. *PloS One* 2014;9:e85560. doi:10.1371/journal.pone.0085560.
- [22] Mooney LM, Rouse EJ, Herr HM. Autonomous exoskeleton reduces metabolic cost of human walking. *J Neuroengineering Rehabil* 2014;11:151. doi:10.1186/1743-0003-11-151.
- [23] Wolpert DM, Kawato M. Multiple paired forward and inverse models for motor control. *Neural Netw* 1998;11:1317–29.
- [24] Hogle LF. Enhancement Technologies and the Body. *Annu Rev Anthropol* 2005;34:695–716. doi:10.1146/annurev.anthro.33.070203.144020.
- [25] Camporesi S. Oscar Pistorius, enhancement and post-humans. *J Med Ethics* 2008;34:639. doi:10.1136/jme.2008.026674.